

Data taken at the 2015 beam test P. Colas



March 1-14, 2015


D. Bhattacharya, P. Colas, F. Couderc, R. Diener, M. Dixit, K. Fujii, S. Ganjour, J. Kaminski, R. Mehdiyev, B. Tuchming (Remotely: D. Attié et al.)




Arrival at DESY March 1, 2015 evening.

Several incidents delayed the start of the test: disparition of our 30 kW LV supply, Stage often stuck, lack of CO2 in TRACI, one module broken while mounting (S8G). Replace by S10X.

Then start gas on March 5 and 2PCO2 cooling. Take data.







On Saturday March 7, frost was found on the pipes: lack of CO2, have to wait until March 8.

Decided to dismount S5D and replace it by S3B.


BD2 limited to 370 V.






Frost deposited on the pipes (-10 °C)







Beam positions

