


Andreas Bamberger, Uwe Renz, Andreas Zwerger

*renz@physik.uni-freiburg.de

TPC - basic operation and pixelated readout





- Charged particle ionizes the chamber gas
- The primary charge drifts along the electric field E towards the end plates
- Magnetic field parallel to E reduces transverse diffusion and allows measurement of the particle momentum
- At the end plate the primary ionization is amplified
- Readout of the signal
- Micro Pattern Gas Detectors (MPGD): high resolution readout and high rates
- Pixelated readout offers high resolution Digital Bubble Chamber
 - A TPC at the ILC needs excellent momentum resolution ⇒ GEM-TimePix for end plate readout

region of interest for this talk


GEM - Gas Electron Multiplier


F. Sauli, Nucl. Instrum. Methods A386(1997)531 F. Sauli, http://www.cern.ch/GDD

Thin, metal coated polymer foil. Very high density of chemical pierced holes

Principle:

- 2 layers Cu each 5µm thick, separated from each other by 50µm Kapton.
- Conical etched holes largest Ø70µm, diagonal distance of holes 140µm.






Advantages of Triple-GEM-setup

- > Gas gain up tp 10^5 in ArCO₂ easily achievable
- > Minimizing the backdrift of positive ions into the drift volume (O(10⁻²))


w

> Encapsulated region of amplification

TimePix properties



DESY Test Beam June 2007



GEM:

10 * 10 cm²



e⁻ beam DESY II

TimePix-GEM-setup

Two different types of GEMs testeted:

trigger (scintill.) & Si-telescope

> Standard 100x100mm² GEMs mit 140µm Lochabstand

small GEMs **24x28mm**² with a fine pitch of **50µm**

6-Oct-08, Amsterdam

GEM-stack with TimePix:

• Shown for fine pitched (small) GEMs

6mm drift volume


•≈ 1.8mm distance between transfer and induction gaps


Data taking properties and reconstruction

- Two independent data streams: TimePix data (Windows based readout) and SiTel.
- SiTel allows the determination of the drift distance with ~50µm precision
- Cleaning up of multiple tracks by
 - double hits in the SiTel
 - in projection along the track: secondary maxi-mum signals another track
- Every other pixel in a row is set to the TOT property, and TIME property in between
- All over the surface ordered in a checker board fashion
- Reconstruction is done by mapping ¹/₂ of all pixels onto 256x256 matrix. Missing neighbours in TOT or TIME data are thereby interpolated




тот	TIME	тот
TIME	тот	TIME
тот	TIME	тот







Special case for tracks close to the first GEM

- The clusters are arranged in a line along the track and produce overlapping charges
- It depends on the direction considered for the precise mapping: rectangular to the track it may stay precise and reveal the hole structure
- the position of the centroids along the track is not mapping the cluster position because of overlapping charge distributions








Check the angular dependance of the structured signal

signal disappears with

- rotation of the coordinate system with respect to the TimePix frame a <-9 and a > 9 mrad
- "fortuitous" peak at a near zero (GEM almost coincides with the chip's orientation)
- the signal strongly depends on the drift distance z

Reason: the disappearance is due to the fact that the transverse diffusion smears out the structure $r_{diff} > d_{hole}$ 1 mm







Discussion of a possible bias for the determination of the point resolution



- does the orientation of the track matter with respect GEM for the point resolution given by the centroids of the clusters?
- Eventually yes, if most of the tracks would run parallel to the holes
- however most tracks are inclined a_{track} ~ (22±6) mrad i.e. far from zero




Resolution of standard GEMs




6-Oct-08, Amsterdam

small GEMs


GEM 50um pitch, triangular pattern, 28X24mm active area holes 30um in active area 4 holes 3,1mm 4 holes 0,3mm on bonding pads (NOT metalized). Leave thick metal on bonding pads (outside possivation square) passivation on BOTTOM SIDE ONLY






- •Referenced outter hole diameter in copper 30μ m.
- •Inner hole diameter in the Kapton between 17μ m- 21μ m.
- ·Diagonal hole pitch $50\mu m$
 - •Projected in $x \approx 43 \mu m$
 - •Projected in $z \approx 25 \mu m$

Resolution of small GEMs



Data processing for time walk correction




original data for com x 10⁴ column in MM

0

- interpolation of TOT and TIME produces 2 sets of data on each pixel
- for any pixels the time walk can be large
- take maximum in TIME and TOT for the correction. In this case at TOT = 180 the correction amounts to +2.7counts to be close to the asymptote, where all clusters should lie.

Treatment of time walk correction




- •Select a column near the centroid of the cluster
- •Fit 2nd order polynomials to the TIME and TOT distributions, look for maxima
- •Time walk correction done only by the maxima of fits
- •Time correction with an exponential function
- •Time correction with an exponential function

$$t_{corr} = TIME_{\max} + Ae^{-tot_{\max}/B}$$

for this run A = 24, B = 81

After correction of time walk for a 100 MHz clock





Similar results from the 48 MHz data (2006) operation: σ_{time} about 12 ns


Measurements with laser



Installation of a laser test bench for measurements with a small number of (primary) electrons.

N2-laser


Detector with metallized drift cathode. The laser releases single photo-electrons \Rightarrow creation of well defined and seperated clusters



- Study single electron clusters
- Investigate influence from number of primary electrons on single point resolution.
- Possibly study ion backdrift proberties

Chip Post-Processing






Pixel acitvePixel passivated

- Work on technologies for post 21 processing chips
- Optimization of readout granularity

Pixelman Design



Pixelman architecture



All basic functions of the Pixelman Software are hardcoded and cannot be changed by the user. The source code is not available.

Pixelman Interface DLL



- Pixelman is Windows based
- Pixelman is written in C++
 - → one is obliged to use the Microsoft Foundation Classes (MFC) which wrap the Windows®Application Interface (API) into a more userfriendly class architecture
- However today's Windows developer use the .NET framework (CLR). This is "incompatible" with Pixelman.

-Oct-08, Amsterdam is already a kind of old fashioned 23

First Plugins

	Printer for Michael Centrel 8 (Annung 2000) Far (2012), New (Second Description				
Inster of let pulse 0 Const Tet hales freed 0 Const Dear Dear Dear Dear Dear Dear Dear Dear		Prime: Difference Notice: Under Notice: Under	Speed in X 21 Speed in Y 21 Speed in Z XY Reps in X 0 Reset Reps in Y 0 Reps in Z 0	Chip Controls No frie selected Chose Mask Acq. time [s] 0 Number of Acq. 0 (per rep.) Synch With Save files to: 	
Adaptation Canada	About About Soc	MpxServer Version 1.0 Copyright (C) 2006 Ket Server P tten by L. TI	lug-In,	More "matur for step mot	

First plug-in was a kind of "toy" plug-in to gain experience. But be warned - as it turns out it is not so straight forward to implement a Graphical User Interface (GUI) into a DLL. But the Prague people helped and therefore thanks a lot!








Can Linux and Windows be "friends"?


...or the question how can we manage on a reasonable (short) time scale to have a Linux based DAQ steering at least basic Pixelman functions, e.g. Start/stop???








Solution \rightarrow use network tools



Program a plug-in that uses standard network connections and TCP/IP, so that Pixelman software can be operated by a Linux DAQ-system, e.g. EUDAQ 26 6-Oct-08, Amsterdam







6-Oct-08, Amsterdam






- Superb spatial resolution of 20-24 μ m
- Time resolution of 8ns after corrections for time walk
- -> z resolution of ~0.5mm in typical TPC gases
- First successful test of pixel with enlarged pixels of $110 \times 110 \mu m^2$
- Reliable operation of TimePix/GEM setup without any special "protective"
 "@dd-ons"





- Finalizing setup of laser test-bench, which could be interesting also for other EUDET partners
- Developing software tools for integration with other DAQ systems and slow control
- BUT everybody is looking forward to a first test beam at DESY -> so focus man power and give a good show ^(C)



- after rotation by 90° the observed spatial signal from the GEM pitch vanishes
- effect on the resolution is an improvement of ~ 2 μm

