

Processing TB2020 SRS raw data

Oleksandr Borysov
DESY




FCAL SW Meeting
April 15, 2020

APV 25 sampled signal

SRS Run 16
5 GeV,
Energy scan study, SRS + Telescope

Example plots:
APV 0,2 (Master)




Signal shape can be used to reject the noise



Signal Extraction Optimization

Fit signal with time response function of CR-RC filter:




$$V = \frac{e^{-\frac{t-t_0}{\tau}} (t-t_0) P_0}{\tau}$$




Time bin with signal maximum

SRS Run 16
5 GeV,
Energy scan study, SRS + Telescope

Example plots:
APV 2 (Master)





Position of the signal maximum



Position of the signal maximum

Max(max_bin(signal_array))



- Even APVs, (masters) have signal maximum in time bin 6;
- Odd APVs, (slaves) have signal maximum in time bin 5;



Position of the signal maximum


Max(cumulative_signal_array)

- Even APVs, (masters) have signal maximum in time bin 6;
- Odd APVs, (slaves) have signal maximum in time bin 5;




APVs noise run 182


Masters, high gain
















Slaves, low gain



Masters, high gain



Slaves, low gain



NN Training

- MPV: 50 ADC
- Noise: Mean: 15, sigma: 5

[training_sig_sim_land_50_noise_15_t0_3_15k.txt](#)


```
Training Neural Network...  
Iteration 500 | Cost: 1.564154e-01  
Program paused. Press enter to continue.
```

Visualizing Neural Network...

Training Set Accuracy: 98.614286

Cross validation Set Accuracy: 98.300000

Test Set Accuracy: 97.775000



- MPV: 50 ADC
- Noise: Mean: 30, sigma: 10

[training_sig_sim_land_50_noise_30_t0_4_15k.txt](#)


```
Training Neural Network...  
Iteration 500 | Cost: 3.551756e-01  
Program paused. Press enter to continue.
```

Visualizing Neural Network...

Training Set Accuracy: 94.700000

Cross validation Set Accuracy: 92.825000

Test Set Accuracy: 92.425000



- MPV: 180 ADC
- Noise: Mean: 30, sigma: 10

[training_sig_sim_land_180_noise_30_t0_4_15k.txt](#)

```
Training Neural Network...  
Iteration 500 | Cost: 9.489138e-02  
Program paused. Press enter to continue.
```

Visualizing Neural Network...

Training Set Accuracy: 99.985714

Cross validation Set Accuracy: 100.000000

Test Set Accuracy: 100.000000

- MPV: 180 ADC
- Noise: Mean: 60, sigma: 20

[training_sig_sim_land_180_noise_60_t0_4_15k.txt](#)

```
Training Neural Network...  
Iteration 500 | Cost: 1.998111e-01  
Program paused. Press enter to continue.
```

Visualizing Neural Network...




Training Set Accuracy: 99.428571

Cross validation Set Accuracy: 97.825000

Test Set Accuracy: 98.200000

NN performance

training_sig_sim_land_180_noise_30_t0_4_15k.txt



NN performance

training_sig_sim_land_180_noise_60_t0_4_15k.txt

For the threshold at 0.7

Correct predictions:




n/n_data: 14816 / 15000 = 98.7733%

Correct signal:





n/n_data: 7492 / 7558 = 99.1268%

Correct noise:





n/n_data: 7324 / 7442 = 98.4144%



















Run 77 event 10



Run 77 event 10



Run 77 event 10



Preprocessed tree

```
p_recotree->Branch("apv_evt",      &m_apv_evt);
p_recotree->Branch("apv_time_s",  &m_apv_time_s);
p_recotree->Branch("apv_time_us", &m_apv_time_us);
p_recotree->Branch("apv_fecNo",   &p_m_apv_fec);
p_recotree->Branch("apv_id",     &p_m_apv_id);
p_recotree->Branch("apv_ch",     &p_m_apv_ch);

p_recotree->Branch("apv_signal_maxbin", &p_m_apv_signal_maxbin); // maximum value of the bins in signal range
p_recotree->Branch("apv_signal_bint1",  &p_m_apv_signal_bint1); // value of time bin 10
p_recotree->Branch("apv_signal_bint2",  &p_m_apv_signal_bint2); // value of time bin 11

p_recotree->Branch("apv_max_bin", &p_m_apv_maxbin); // number of the time bin with maximum signal
p_recotree->Branch("apv_bint1",   &p_m_apv_bint1); // number of the fixed time bin for this apv and this run

p_recotree->Branch("apv_signal_maxfit", &p_m_apv_signal_maxfit); // maximum of the fit with RC-CR response function
p_recotree->Branch("apv_fit_t0",       &p_m_apv_fit_t0); // t0 of the fit with RC-CR response function
p_recotree->Branch("apv_fit_tau",     &p_m_apv_fit_tau); // tau of the fit with RC-CR response function
p_recotree->Branch("apv_fit_chi2",    &p_m_apv_fit_chi2); // chi2 of the fit with RC-CR response function
p_recotree->Branch("apv_nn_output",   &p_m_apv_nn_output); // neural network output

p_recotree->Branch("apv_cm",         &p_m_apv_cm); // commomn mode noise
```

```
.....
std::vector <UInt_t>    m_apv_ch;
std::vector <Double_t> m_apv_signal_maxbin;
.....
```

<https://www.desy.de/~oborysov/tb2020/index.html>

[Process_apv_data.h](#)

[Process_apv_data.C](#)

} SRS "raw"
tree

[APV_Reco_Tree.h](#)

[APV_Reco_Tree.C](#)

[ReadRecoTest.C](#)

} Preprocessed
tree

[info.txt](#)

Short 15
instructions

SRS “raw” tree

```
p_rawtree->SetBranchAddress("apv_evt",    p_m_apv_evt);
p_rawtree->SetBranchAddress("time_s",    p_m_time_s);
p_rawtree->SetBranchAddress("time_us",   p_m_time_us);
p_rawtree->SetBranchAddress("apv_fecNo", &p_m_apv_fec);
p_rawtree->SetBranchAddress("apv_id",    &p_m_apv_id);
p_rawtree->SetBranchAddress("apv_ch",    &p_m_apv_ch);
p_rawtree->SetBranchAddress("mm_id",     &p_m_mm_id);
p_rawtree->SetBranchAddress("mm_readout", &p_m_mm_readout);
p_rawtree->SetBranchAddress("mm_strip",  &p_m_mm_strip);

p_rawtree->SetBranchAddress("apv_q",      &p_m_apv_q);
p_rawtree->SetBranchAddress("apv_presamples", &p_m_apv_presamples);
```


Summary

- SRS TB2020 and TB2019 data were preprocessed using the procedure similar to TB2016.
- Several signal estimations are available:
 - Value from the expected maximum time bin (6 for master and 5 for slave);
 - Maximum within the range around expected value (time bins 1-15);
 - Amplitude of the fit with CR-RC response function;
- Output of neural network trained separately for different gain and noise level;
- Parameter of CR-RC response function: t_0 , τ and Chi^2 .
- There is significant difference in signal shape between modified and non-modified APV. Can be taken into account for NN training when confirmed and understood.
- Deconvolution can be also used for signal estimation.